Ph: 02 9417 6660 E: info@alphabond.com.au W: www.alphabond.com.au

ALPHABOND dental

JENSEN DENTAL

Application example

Molar

by ZTM Axel Gütges

www.miyoworld.eu

Pre Crystallization

850 MiYO

850 MiYO

Application example

Molar





Note:
850 MiYO Color diagram for illustrative purposes only. Any 850 MiYO Color may be used based on your case needs.

1 Preparation

- Prepare the lithium disilicate frameworks according to the manufacturer's instructions.
- Mix 850 MiYO Color and Glaze Paste thoroughly with a metal free spatula.
- Add a thin layer of 850 MiYO Glaze Paste before you start coloring.

2 Application 850 MiYO Color / Glaze

- 850 MiYO Colors can be used individually or in combination.
- Apply the materials to the framework according to their intended use.
- The entire surface must be covered with 850 MiYO Colors or 850 MiYO Glaze Paste.

3 Firing

Use the pre-programmed standard or speed cycles for crystallization or the pre-programmed correction cycles for post crystallization.
If using a furnace that does not have pre-programmed cycles, follow manufacturer recommended firing parameters for crystallization.

Finished 850 MiYO work by ZTM Axel Gütges

EN

3101010525